



Product Technical Information

Effective: July 1, 1999

Re: Nickel Content

Fixation pins contained in the Agee WristJack® System are made of type 316, implant grade stainless steel. This alloy is composed of approximately 14% NICKEL, among other metals.

For patients who are allergic to nickel, WristJack fixation pins may not be suitable. Both the treating surgeon and the patient should be notified of the nickel content.

We do not sell or know of an alternative pin for these patients. Contact Hand Biomechanics Lab at (800) 522-5778 if you have any questions.